

ANKAphase: software for single-distance phase-retrieval from inline X-ray phase contrast radiographs. Erratum

T. WEITKAMP,^{a*} D. HAAS,^{b,c} D. WEGRZYNEK^d AND A. RACK^{e*}

^a*Synchrotron Soleil, Gif-sur-Yvette, France,* ^b*Institute for Synchrotron Radiation / ANKA Light Source, KIT, Karlsruhe, Germany,* ^c*Hochschule Darmstadt, Darmstadt, Germany,* ^d*International Atomic Energy Agency, Seibersdorf, Austria (present address: University of Science and Technology, Kraków, Poland), and* ^e*European Synchrotron Radiation Facility, Grenoble, France.*

E-mail: weitkamp@synchrotron-soleil.fr, arack@snafu.de

(Received 0 XXXXXXXX 0000; accepted 0 XXXXXXXX 0000)

phase retrieval; X-ray phase contrast; synchrotron radiation; microtomography; X-ray imaging; Java; ImageJ

There was a mistake in Equation (6) of our paper. The equation should read

$$\phi(x, y) = \frac{\delta}{2\beta} \ln \left(\mathcal{F}^{-1} \left\{ \frac{\mathcal{F}[I(x, y)/I_0(x, y)]}{1 + [\lambda z \delta / (4\pi\beta)] (u^2 + v^2)} \right\} \right) . \quad (6)$$

References

Weitkamp, T., Haas, D., Wegrzynek, D. & Rack, A. (2011). *J. Synchrotron Rad.* **18**, 617–629.

Synopsis

Equation (6) in the paper by Weitkamp *et al.* [(2011), *J. Synchrotron Rad.* **18**, 617–629] is corrected.
